

United States Department of Agriculture  
Center for Veterinary Biologics  
Testing Protocol

## SAM 502

Supplemental Assay Method for the Determination of  
Residual Moisture in Veterinary Biologics Products

Date: October 20, 2004  
Supersedes: New  
Number: TCSAM0502.01  
Standard Requirement: 9 CFR 113.29  
Contact Person: P. Frank Ross, (515) 663-7542

## Approvals:

/s/P. Frank Ross Date: 20Oct04  
P. Frank Ross, Head  
Chemistry and Analytical Services Section

/s/Byron E. Rippke Date: 22Oct04  
Richard E. Hill, Jr., Director  
Policy, Evaluation, and Licensing  
Center for Veterinary Biologics

/s/ Rebecca L.W. Hyde Date: 22OCT04  
Rebecca L.W. Hyde, Section Leader  
Quality Assurance  
Center for Veterinary Biologics

United States Department of Agriculture  
Animal and Plant Health Inspection Service  
P. O. Box 844  
Ames, IA 50010

Mention of trademark or proprietary product does not constitute a guarantee or warranty of the product by USDA and does not imply its approval to the exclusion of other products that may be suitable.

Supplemental Assay Method for the Determination of Residual Moisture  
in Veterinary Biologics Products

Table of Contents

1. Introduction
2. Materials
  - 2.1 Equipment
3. Preparation for the test
  - 3.1 Personnel qualifications/training
  - 3.2 Preparation of equipment/instrumentation
  - 3.3 Preparation of reagents/control procedures
  - 3.4 Preparation of the sample
4. Performance of the test
5. Interpretation of the test results
6. Report of the test results
7. References

**Supplemental Assay Method for the Determination of Residual Moisture  
in Veterinary Biologics Products**

## **1. Introduction**

Freeze-dried veterinary vaccines always contain some water, commonly known as residual moisture (RM). It is important to determine the level of RM in final products, since a satisfactory test gives assurance of an adequate shelf life and that the manufacturer's freeze-dry cycle was properly controlled. The RM test should confirm that moisture level is consistently within the manufacturer's specification.

This Supplemental Assay Method (SAM) describes the test procedure for residual moisture testing for final product testing of freeze dried new veterinary vaccines.

Residual moisture is determined by the gravimetric method as follows: Residual moisture is driven from the test product by heating under vacuum. The residual moisture content (as per cent) of the test product is calculated based the product weight loss during the drying cycle.

## **2. Materials**

### **2.1 Equipment**

**2.1.1** Cylindrical weighing bottles--individually numbered with airtight glass stoppers.

**2.1.2** Vacuum oven--equipped with validated thermometer and thermostat. A suitable air-drying device must be attached to the inlet valve.

**2.1.3** Balance--capable of readability to 0.1 mg (rated precision  $\pm 0.1$  mg).

**2.1.4** Desiccator--with phosphorus pentoxide, silica gel or equivalent

## **3. Preparation for the test**

**3.1 Personnel training**--No specific training is required. Individual should have working knowledge of laboratory equipment listed in **Section 2**.

### **3.2 Preparation for the test - Environment**

**Supplemental Assay Method for the Determination of Residual Moisture  
in Veterinary Biologics Products**

Conduct all operations in an environment with a relative humidity less than 45%.

**3.3 Preparation for the test - Weighing bottles**

Label the weighing bottle for sample(s). Thoroughly clean weighing bottles. Place stopper at an angle on top of bottle and dry for a minimum of 30 minutes at  $60^{\circ} \pm 3^{\circ}\text{C}$  under vacuum ( $<2.5$  kPa). While hot, immediately transfer bottles and stoppers into a desiccator. Allow to cool to room temperature, close stopper, weigh and record the weights as "A". Return bottles to desiccator.

**3.4 Preparation of the sample** - Retain sample, in original airtight containers at room temperature until use. Do not break the seal until ready to proceed.

**4. Performance of the test**

**4.1** Break sample container seal. Using a spatula, break up desiccated product and rapidly transfer (minimum of 100 mg or the amount required for a precise determination at the lower limit, use more than one vial for single dose products if needed) to a previously weighed bottle. Close stopper and immediately weigh. Record the weight as "B".

**4.2** Place the bottle with the stopper at an angle in the vacuum oven. Set vacuum to  $<2.5$  kPa and the temperature to  $60^{\circ} \pm 3^{\circ}\text{C}$ .

**4.3** After a minimum of 3 hrs, turn off the vacuum pump and bleed dry air into the oven until the pressure inside of the oven is equalized with the atmosphere.

**4.4** While the bottle is still warm, stopper bottle and transfer to desiccator, and allow to cool to room temperature (for a minimum of two hours or a time validated to yield a constant weight). Weigh, and record the weight as "C".

Supplemental Assay Method for the Determination of Residual Moisture  
in Veterinary Biologics Products

**5. Interpretation of results**

Calculate the residual moisture (%) as

$$\% \text{ moisture} = ((B - C) / (B - A)) \times (100)$$

**A** is tare weight of bottle.

**B** minus **A** is weight of sample before assay.

**B** minus **C** is weight equivalent to residual moisture of sample.

**6. Report of test results**

Test results are reported according to current procedures.

**7. References**

**7.1** Code of Federal Regulations, Title 9, Parts 113.29

**7.2** Testing of Residual Moisture, VICH GL26, Final,  
April, 2002